



Petrifilm™

Created by Authorized Personnel: Lola Johnson, 2019-01-11

Manufacture Date: 2019-01-04

Expiration Date: 2020-07-04

Product Manufacturing Certificate

Certificate of Analysis

Product: 6407, 6417 or 6445 3M™ Petrifilm™ Yeast and Mold Count Plates
 Batch: 333SJW
 Stock Number: 70-2005-6124-2 or 70-2005-9015-9 or 70-2007-7904-2
 ERP Number: 7100039379 or 7100039423 or 7100047854

Organism Tested	Minimum Growth	Result
<i>Aspergillus niger</i> , M6	Productivity Ratio \geq 0.7	Pass
<i>Paecilomyces</i> species, M10	Productivity Ratio \geq 0.7	Pass
<i>Botrytis</i> species, M97	Productivity Ratio \geq 0.7	Pass
<i>Hansenula anomala</i> , Y28	Productivity Ratio \geq 0.7	Pass
<i>Saccharomyces cerevisiae</i> , Y41	Productivity Ratio \geq 0.7	Pass
<i>Enterococcus faecalis</i>	No Growth	Pass
<i>Acinetobacter</i> species	No Growth	Pass

This material complies with the 3M specifications for this product construction, and applicable criteria for routine quality control and microbiological performance. 3M Brookings is certified to ISO 9001 through an independent agency and is an FDA registered drug and device site.

3M Health Care PO Box 5227
 Brookings, SD 57006-5227 Phone: 1-800-328-1671

3M and Petrifilm are trademarks of 3M. Please recycle.
 Printed in USA © 3M 2017. All rights reserved.

Version 6

YM
 Yeast and Mold Count Plate