



Petrifilm™

Created by Authorized Personnel: Lola Johnson, 2019-02-21

Manufacture Date: 2019-02-11

Expiration Date: 2020-08-11

Product Manufacturing Certificate

Certificate of Analysis

Product: 3M™ Petrifilm™ Rapid Yeast and Mold Count Plates 6475 or 6477
Batch: 333YNT
Stock Number: 70-2007-7579-2 or 70-2007-7580-0
ERP Number: 7100039447 or 7100040141

Organism Tested	Growth Specification	Result
<i>Geotrichum candidum</i> ATCC 34614	Productivity Ratio \geq 0.5	Pass
<i>Candida albicans</i> ATCC 10231	Productivity Ratio \geq 0.5	Pass
<i>Enterococcus faecalis</i> ATCC 14506	No Growth	Pass
<i>Escherichia coli</i> ATCC 25922	No Growth	Pass

This material complies with the 3M specifications for this product construction, and applicable criteria for routine quality control and microbiological performance. 3M Brookings is certified to ISO 9001 through an independent agency.

3M Health Care
3M Center, Building 275-5W-05
St. Paul, MN 55144-1000 Phone: 1-800-328-1671

3M and Petrifilm are trademarks of 3M. Please recycle.
Printed in USA © 3M 2017. All rights reserved.

Version 2

RYM
Rapid Yeast and Mold Count Plate

