



Food Safety

3M Center

St. Paul, MN 55144-1000

www.3M.com/foodsafety

Certificate of Analysis

Product Identification

| | |
|----------------------|--|
| 3M Catalog Number | L25EGG |
| Product Description | 3M™ Egg White Protein Rapid Kit - 25 Tests per Kit, 1 Kit per Case |
| 3M Stock Number | 70-2011-7549-7 |
| 3M Kit Lot Number | RD00410 |
| 3M Buffer Lot Number | RD00410 |
| Kit Expiration Date | 4/12/2020 |

Performance Testing*

Negative Controls:


| Sample | Expected Results | Pass/Fail |
|----------------|---------------------|-----------|
| Matrix Testing | Negative, ≥ 24 | Pass |

Positive Controls:

| Sample | Expected Results | Pass/Fail |
|----------------|---------------------|-----------|
| Matrix Testing | Positive, ≥ 24 | Pass |

*A minimum of four matrices are used for performance testing. All results are interpreted by eye, holding the strip an arm's length away. If the control, hook and test lines are visible the test is considered positive. If only the control and hook lines are visible the test is considered negative. If either the control or hook lines aren't visible the test is considered invalid. Positive controls are spiked at 0.5ppm. This letter certifies that the product described above has been tested and found to comply with 3M's product performance specifications applicable to subject device at time of manufacture. Visit the 3M Food Safety website, www.3m.com/foodsafety for additional information.

Prepared By:

| | |
|---------|--|
| Signed: |  |
| Name: | Margaret Haws QA/QC Representative |
| Date: | 5/22/2018 |