



Petrifilm™

Created by Authorized Personnel: Lola Johnson, 2019-05-21

Manufacture Date: 2019-04-28

Expiration Date: 2020-10-26

Product Manufacturing Certificate

Certificate of Analysis

Product: 3M™ Petrifilm™ Aqua Coliform Count Plates 6457 or 6458
 Batch: 334DL8
 Stock Number: 70-2007-6474-7 or 70-2007-6475-4
 ERP Number: 7100039445 or 7100039446

Organism Tested	Minimum Growth	Result
<i>Escherichia coli</i> ATCC 51813	≥-3.0* with gas	Pass
<i>Enterobacter amnigenus</i> ATCC 51816	≥-3.0* with gas	Pass
<i>Escherichia coli</i> ATCC 25922	Productivity Ratio≥0.5	Pass
<i>Staphylococcus aureus</i> ATCC 6538	No Growth	Pass
<i>Enterococcus faecalis</i> ATCC 29212	No Growth	Pass
<i>Pseudomonas aeruginosa</i> ATCC 27853	No or Atypical Growth	Pass

*Expressed as the number of standard deviations away from the average count on standard agar medium used in the water industry

This material complies with the 3M specifications for this product construction, and applicable criteria for routine quality control and microbiological performance. 3M Brookings is certified to ISO 9001 through an independent agency.

3M Health Care
 3M Center, Building 275-5W-05
 St. Paul, MN 55144-1000 Phone: 1-800-328-1671

3M and Petrifilm are trademarks of 3M. Please recycle.
 Printed in USA © 3M 2017. All rights reserved.

Version 4

AQCC
 Aqua Coliform Count Plate

