



Petrifilm™

Created by Authorized Personnel: Aubrey Poling, 2019-06-19

Manufacture Date: 2019-05-12

Expiration Date: 2020-11-09

Product Manufacturing Certificate

Certificate of Analysis

Product: 6407, 6417 or 6445 3M™ Petrifilm™ Yeast and Mold Count Plates
 Batch: 334FHR
 Stock Number: 70-2005-6124-2 or 70-2005-9015-9 or 70-2007-7904-2
 ERP Number: 7100039379 or 7100039423 or 7100047854

| Organism Tested | Minimum Growth | Result |
|---------------------------------------|-------------------------------|--------|
| <i>Aspergillus niger</i> , M6 | Productivity Ratio \geq 0.7 | Pass |
| <i>Paecilomyces</i> species, M10 | Productivity Ratio \geq 0.7 | Pass |
| <i>Botrytis</i> species, M97 | Productivity Ratio \geq 0.7 | Pass |
| <i>Hansenula anomala</i> , Y28 | Productivity Ratio \geq 0.7 | Pass |
| <i>Saccharomyces cerevisiae</i> , Y41 | Productivity Ratio \geq 0.7 | Pass |
| <i>Enterococcus faecalis</i> | No Growth | Pass |
| <i>Acinetobacter</i> species | No Growth | Pass |

This material complies with the 3M specifications for this product construction, and applicable criteria for routine quality control and microbiological performance. 3M Brookings is certified to ISO 9001 through and independent agency.

3M Health Care PO Box 5227
 Brookings, SD 57006-5227 Phone: 1-800-328-1671

3M and Petrifilm are trademarks of 3M. Please recycle.
 Printed in USA © 3M 2017. All rights reserved.

Version 6

YM

Yeast and Mold Count Plate