



# Petrifilm™

Created by Authorized Personnel: Lola Johnson, 2019-08-08

Manufacture Date: 2019-08-03

Expiration Date: 2021-01-31

## Product Manufacturing Certificate

## Certificate of Analysis

Product: 3M™ Petrifilm™ *E. coli*/Coliform Count Plates 6404 or 6414 or 6444  
 Batch: 334YWK  
 Stock Number: 70-2005-7213-2 or 70-2005-9014-2 or 70-2007-7073-6  
 ERP Number: 7100039312 or 7100039397 or 7100047857

Organism Tested	Growth Specification	Result
<i>Enterococcus faecalis</i> ATCC 14506	No Growth	Pass
<i>Enterobacter amnigenus</i> ATCC 51818	≥-3.0*, red with gas	Pass
<i>Escherichia coli</i> ATCC 51813	≥-3.0*, blue with gas	Pass
<i>Hafnia alvei</i> ATCC 51815	≥-3.0*, red with gas	Pass
<i>Klebsiella oxytoca</i> ATCC 51817	≥-3.0*, red with gas	Pass
<i>Salmonella typhimurium</i> ATCC 51812	≥-3.0*, red with gas	Pass

\*Expressed as the number of standard deviations away from the average count on standard agar medium.

### ISO 11133:2014 Performance Testing

Organism Tested	Growth Specification	Result
<i>Escherichia coli</i> ATCC 25922 (WDCM 00013)	Productivity Ratio≥0.5	Pass
<i>Escherichia coli</i> ATCC NCTC 13216 (WDCM 00202)	Productivity Ratio≥0.5	Pass
<i>Escherichia coli</i> ATCC 8739 (WDCM 00012)	Productivity Ratio≥0.5	Pass
<i>Citrobacter freundii</i> 43864 (WDCM 00007)	Atypical Growth	Pass
<i>Pseudomonas aeruginosa</i> 27853 (WDCM 00025)	Atypical Growth	Pass
<i>Enterococcus faecalis</i> ATCC 29212 (WDCM 00087)	No Growth	Pass

This material complies with the 3M specifications for this product construction, and applicable criteria for routine quality control and microbiological performance. 3M Brookings is certified to ISO 9001 through an independent agency.

3M Health Care  
3M Center, Building 275-5W-05  
St. Paul, MN 55144-1000 Phone: 1-800-328-1671

3M and Petrifilm are trademarks of 3M. Please recycle.  
Printed in USA © 3M 2017. All rights reserved.

Version 7

# EC

## *E. coli*/Coliform Count Plate