



Certificate of Analysis

Product Name: 3M Clean-Trace Surface ATP

SAP/ERP Material Number: 7100067519

3M Stock Number: GH620522827

Kit Batch Number: ...19483217...

Date of Manufacture: ...25 Nov 2019...

Date of Expiry: ...25 Sep 2020...

| Test | Temp | Inst Serial Number | Specification | Test Result |
|-------------------|------|--------------------|----------------------------------|-------------|
| Device Background | 20°C | 105 | Mean Bk \leq 25 RLU | Pass |
| Device Activity | 20°C | 105 | Activity Mean 7109 – 7972 RLU | Pass |

This is to certify that the product described above has been manufactured in accordance with 3M specifications.

AUTHORISED: (Quality Control)

DATE: ...27 Nov 2019...

3M UK PLC

The Science Park
Bridgend, CF31 3NA
Tel +44 (0) 1656 641400
Fax +44 (0) 1656 768835
www.3m.com/foodsafety