



Petrifilm™

Created by Authorized Personnel: Lola Johnson, 2019-12-09

Manufacture Date: 2019-11-26

Expiration Date: 2021-05-26

Product Manufacturing Certificate

Certificate of Analysis

Product: 3M™ Petrifilm™ Rapid Aerobic Count Plates 6478 or 6479
Batch: 3366LC
Stock Number: 70-2007-8453-9 or 70-2007-8454-7
ERP Number: 7100039448 or 7100039449

ISO 11133:2014 Performance Testing

Organism Tested	Growth Specification	Result
<i>Escherichia coli</i> ATCC 25922 (WDCM 00013)	Productivity Ratio \geq 0.7	Pass
<i>Escherichia coli</i> ATCC 8739 (WDCM 00012)	Productivity Ratio \geq 0.7	Pass
<i>Staphylococcus aureus</i> ATCC 25923 (WDCM 00034)	Productivity Ratio \geq 0.7	Pass
<i>Bacillus subtilis</i> subsp. <i>spizizenii</i> ATCC 6633 (WDCM 00003)	Productivity Ratio \geq 0.7	Pass

This material complies with the 3M specifications for this product construction, and applicable criteria for routine quality control and microbiological performance of ISO 11133. 3M Brookings is certified to ISO 9001 through an independent agency.

3M Health Care
3M Center, Building 275-5W-05
St. Paul, MN 55144-1000 Phone: 1-800-328-1671

3M and Petrifilm are trademarks of 3M. Please recycle.
Printed in USA © 3M 2017. All rights reserved.

Version 3

RAC
Rapid Aerobic Count Plate