



# Petrifilm™

Created by Authorized Personnel: Małgorzata Stec, 2020-01-16

Manufacture Date: 2020-01-02

Expiration Date: 2021-07-02

## Product Manufacturing Certificate

## Certificate of Analysis

Product: 3M™ Petrifilm™ Staph Express Count Plates 6446, 6490 or 6491  
 Batch: 418320001A  
 Stock Number: 70-2007-9229-2 or 70-2007-9230-0 or 70-2007-9228-4  
 ERP Number: 7100090291 or 7100090292 or 7100090307

| Organism Tested                              | Growth Specification          | Result |
|----------------------------------------------|-------------------------------|--------|
| <i>Staphylococcus aureus</i> ATCC 49476      | Productivity Ratio $\geq$ 0.5 | Pass   |
| <i>Staphylococcus epidermidis</i> ATCC 14990 | No or Atypical Growth         | Pass   |
| <i>Bacillus atrophaeus</i> ATCC 9372         | No or Atypical Growth         | Pass   |

### ISO 11133:2014 Performance Testing

| Organism Tested                                             | Growth Specification          | Result |
|-------------------------------------------------------------|-------------------------------|--------|
| <i>Staphylococcus aureus</i> ATCC 6538 (WDCM 00032)         | Productivity Ratio $\geq$ 0.5 | Pass   |
| <i>Staphylococcus aureus</i> ATCC 25923 (WDCM 00034)        | Productivity Ratio $\geq$ 0.5 | Pass   |
| <i>Escherichia coli</i> ATCC 25922 (WDCM 00013)             | No Growth                     | Pass   |
| <i>Escherichia coli</i> ATCC 8739 (WDCM 00012)              | No Growth                     | Pass   |
| <i>Staphylococcus epidermidis</i> ATCC 12228 (WDCM 00036)   | No or Atypical Growth         | Pass   |
| <i>Staphylococcus saprophyticus</i> ATCC 15305 (WDCM 00159) | No or Atypical Growth         | Pass   |

This material complies with the 3M specifications for this product construction, and applicable criteria for routine quality control and microbiological performance of ISO 11133. 3M Wroclaw is certified to ISO 9001 through an independent agency.

3M Health Care  
 3M Center, Building 275-5W-05  
 St. Paul, MN 55144-1000 Phone: 1-800-328-1671

3M and Petrifilm are trademarks of 3M. Please recycle.  
 Printed in USA © 3M 2017. All rights reserved.

Version 11

**STX**  
 Staph Express Count Plate