

# 3M Health Care

3M Center  
St. Paul, MN 55144-1000

## Certificate of Analysis

### Product Identification

3M Catalog Number	3MBPW500
Product Description	3M 500g Buffered Peptone Water ISO
3M Stock Number	70-2011-7372-4
3M Lot Number	A200601
Expiration Date	January 6, 2025

### Specifications of product:


**Appearance:** Clear, yellowish liquid  
**pH (acceptable range: 6.8-7.2):** 6.98

### Biological Performance Testing:

Test Strain	Analysis	Disposition
<b>Acceptable Range</b>	<b>10-100 CFU</b>	<b>Pass</b>
Enterococcus faecalis ATCC 33186	89	Pass
Escherichia coli ATCC 25922	72	Pass
Pseudomonas aeruginosa ATCC 27853	74	Pass
Salmonella enteritidis ATCC 13076	99	Pass
Salmonella typhimurium ATCC 14028	59	Pass

This letter certifies that the product described above has been tested and found to comply with 3M's product performance specifications applicable to subject device at time of manufacture.

### Prepared By:

Signed:	
Name:	Patrick McMonagle QA/QC Representative
Date:	3/2/2020