



Petrifilm™

Created by Authorized Personnel: Lola Johnson, 2020-01-20

Manufacture Date: 2020-01-10

Expiration Date: 2021-07-10

Product Manufacturing Certificate

Certificate of Analysis

Product: 3M™ Petrifilm™ *Salmonella* Express Plate 6536 or 6537
Batch: 336JJH
Stock Number: 70-2007-6998-5 or 70-2007-6999-3
ERP Number: 7100039574 or 7100039575

Organism Tested	Minimum Growth	Result
<i>Salmonella enterica</i> subsp <i>enterica</i> Typhimurium ATCC 14028	Typical Growth	Pass
<i>Salmonella enterica</i> subsp <i>enterica</i> Enteritidis ATCC 49223	Typical Growth	Pass
<i>Citrobacter freundii</i> FSD 548	Atypical Growth	Pass
<i>Escherichia coli</i> FSD 723	Atypical Growth	Pass

This material complies with the 3M specifications for this product construction, and applicable criteria for routine quality control and microbiological performance. 3M Brookings is certified to ISO 9001 through an independent agency.

3M Health Care
3M Center, Building 275-5W-01
St. Paul, MN 55144-1000 Phone: 1-800-328-1671

3M and Petrifilm are trademarks of 3M. Please recycle.
Printed in USA © 3M 2017. All rights reserved.

Version 2

SALX
Salmonella Express Plate