

3M Health Care

3M Center
St. Paul, MN 55144-1000

Certificate of Analysis

Product Identification

3M Catalog Number	3MBPW500
Product Description	3M 500g Buffered Peptone Water ISO
3M Stock Number	70-2011-7372-4
3M Lot Number	C200203
Expiration Date	March 10, 2025

Specifications of product:


Appearance: Yellow
pH (acceptable range: 6.8-7.2): 6.99

Biological Performance Testing:

Test Strain	Analysis	Disposition
Acceptable Range	10-100 CFU	Pass
Enterococcus faecalis ATCC 33186	71 CFU	Pass
Escherichia coli ATCC 25922	64 CFU	Pass
Pseudomonas aeruginosa ATCC 27853	75 CFU	Pass
Salmonella enteritidis ATCC 13076	72 CFU	Pass
Salmonella typhimurium ATCC 14028	74 CFU	Pass

This letter certifies that the product described above has been tested and found to comply with 3M's product performance specifications applicable to subject device at time of manufacture.

Prepared By:

Signed:	
Name:	Patrick McMonagle QA/QC Representative
Date:	4/17/2020