



Petrifilm™

Created by Authorized Personnel: Małgorzata Stec, 2019-12-18

Manufacture Date: 2019-12-12

Expiration Date: 2021-06-11

Product Manufacturing Certificate

Certificate of Analysis

Product: 3M™ Petrifilm™ *Enterobacteriaceae* Count Plates 6420 or 6421
 Batch: 418319346A
 Stock Number: UU-0081-9643-6 or UU-0081-9646-9
 ERP Number: 7100126818 or 7100126807

Organism Tested	Growth Specification	Result
<i>Enterobacter amnigenus</i> ATCC 51818	≥-3.0*, gas and acid	Pass
<i>Enterobacter amnigenus</i> ATCC 51816	≥-3.0*, gas and acid	Pass
<i>Salmonella enterica</i> subsp. <i>enterica</i> Typhimurium ATCC 51812	≥-3.0*, gas and acid	Pass
<i>Pseudomonas aeruginosa</i> ATCC 35554	Red Colonies without gas or acid	Pass
<i>Enterococcus faecalis</i> ATCC 14506	No Growth	Pass

*Expressed as the number of standard deviations away from the average count on standard agar medium.

ISO 11133:2014 Performance Testing

Organism Tested	Growth Specification	Result
<i>Escherichia coli</i> ATCC 25922 (WDCM 00013)	Productivity Ratio≥0.5	Pass
<i>Escherichia coli</i> ATCC 8739 (WDCM 00012)	Productivity Ratio≥0.5	Pass
<i>Salmonella enterica</i> subsp. <i>enterica</i> Typhimurium ATCC 14028 (WDCM 00031)	Productivity Ratio≥0.5	Pass
<i>Enterococcus faecalis</i> ATCC 29212 (WDCM 00087)	No Growth	Pass

This material complies with the 3M specifications for this product construction, and applicable criteria for routine quality control and microbiological performance of ISO 11133. 3M Wroclaw is certified to ISO 9001 through an independent agency.

3M Health Care
3M Center, Building 275-5W-05
St. Paul, MN 55144-1000 Phone: 1-800-328-1671

3M and Petrifilm are trademarks of 3M. Please recycle.
Printed in USA © 3M 2017. All rights reserved.

Version 8

EB

Enterobacteriaceae Count Plate