

3M Health Care

3M Center
St. Paul, MN 55144-1000

Certificate of Analysis & Sterility

Product Identification


3M Catalog Number	6432
Product Description	3M™ Quick Swab
3M Stock Number	70200684887
3M Lot Number	10320
Expiration Date Code (MM-DD-YYYY)	08/07/2021

Test Data:

Test	Minimum	Maximum	Result
pH	6.8	7.4	7.36
Gamma Irradiation	25.0 kGy	35.0 kGy	25.8-31.8 kGy

This letter Certifies that the product listed on this certificate has been manufactured in accordance with 3M Health Care specifications.

Prepared By:

Signed:	
Name:	Patrick McMonagle QA/QC Representative
Date:	4/16/2020