



Petrifilm™

Created by Authorized Personnel: Jarosław Matczak, 2020-04-17

Manufacture Date: 2020-04-14

Expiration Date: 2021-10-13

Product Manufacturing Certificate

Certificate of Analysis

Product: 3M™ Petrifilm™ Environmental *Listeria* Plates 6447 or 6448
Batch: 418320105A
Stock Number: 70-2007-9226-8 or 70-2007-9227-6
ERP Number: 7100090324 or 7100090323

Organism Tested	Growth Specification	Result
<i>Listeria monocytogenes</i> SLR-22	Typical <i>Listeria</i> Growth	Pass
<i>Listeria innocua</i> ATCC 33090	Typical <i>Listeria</i> Growth	Pass
<i>Enterococcus faecalis</i> ATCC 14506	No Typical <i>Listeria</i> Growth	Pass
<i>Bacillus atrophaeus (subtilis)</i> ATCC 9372	No Typical <i>Listeria</i> Growth	Pass

This material complies with the 3M specifications for this product construction, and applicable criteria for routine quality control and microbiological performance. 3M Wrocław is certified to ISO 9001 through an independent agency.

3M Health Care
3M Center, Building 275-5W-05
St. Paul, MN 55144-1000 Phone: 1-800-328-1671

3M and Petrifilm are trademarks of 3M. Please recycle.
Printed in USA © 3M 2017. All rights reserved.

Version 5

EL

Environmental *Listeria* Plate