



Certificate of Conformance

Product Name: 3M Clean-Trace Water Plus – Total ATP

SAP/ERP Material Number: 7100002659 3M Stock Number: GH620538112

Kit Batch Number: ...20313213...

Date of Manufacture: ...23 Jul 2020...

Date of Expiry: ...21 Jul 2021...

Test	Temp	Inst Serial Number	Specification	Test Result
Device Background	20°C	105	Mean Bk \leq 20 RLU	Pass
Device Activity	20°C	105	Activity Mean 15000 – 20000 RLU	Pass

This is to certify that the product described above has been manufactured in accordance with 3M specifications.

AUTHORISED: (Quality Control)

DATE: ...28 Jul 2020...

3M UK PLC

The Science Park
Bridgend, CF31 3NA
Tel +44 (0) 1656 641400
Fax +44 (0) 1656 768835
www.3m.com/foodsafety