



Certificate of Analysis

Product Name: 3M Clean-Trace Surface ATP

SAP/ERP Material Number: 7100067519

3M Stock Number: GH620522827

Kit Batch Number: ...20413224...

Date of Manufacture: ...08 Oct 2020...

Date of Expiry: ...08 Aug 2021...

Test	Temp	Inst Serial Number	Specification	Test Result
Device Background	20°C	105	Mean Bk \leq 25 RLU	Pass
Device Activity	20°C	105	Activity Mean 7109 – 7972 RLU	Pass

This is to certify that the product described above has been manufactured in accordance with 3M specifications.

AUTHORISED: (Quality Control)

DATE: ...09 Oct 2020...

3M UK PLC

The Science Park
Bridgend, CF31 3NA
Tel +44 (0) 1656 641400
Fax +44 (0) 1656 768835
www.3m.com/foodsafety