



Petrifilm™

Created by Authorized Personnel: Jarosław Matczak, 2021-01-22

Manufacture Date: 2021-01-06

Expiration Date: 2022-07-07

Product Manufacturing Certificate

Certificate of Analysis

Product: 3M™ Petrifilm™ Staph Express Count Plates 6446, 6490 or 6491
Batch: 418321006C
Stock Number: 70-2007-9229-2 or 70-2007-9230-0 or 70-2007-9228-4
ERP Number: 7100090291 or 7100090292 or 7100090307

Organism Tested	Growth Specification	Result
<i>Staphylococcus aureus</i> ATCC 49476	Productivity Ratio \geq 0.5	Pass
<i>Staphylococcus epidermidis</i> ATCC 14990	No or Atypical Growth	Pass
<i>Bacillus atropheus</i> ATCC 9372	No or Atypical Growth	Pass

ISO 11133:2014 Performance Testing

Organism Tested	Growth Specification	Result
<i>Staphylococcus aureus</i> ATCC 6538 (WDCM 00032)	Productivity Ratio \geq 0.5	Pass
<i>Staphylococcus aureus</i> ATCC 25923 (WDCM 00034)	Productivity Ratio \geq 0.5	Pass
<i>Escherichia coli</i> ATCC 25922 (WDCM 00013)	No Growth	Pass
<i>Escherichia coli</i> ATCC 8739 (WDCM 00012)	No Growth	Pass
<i>Staphylococcus epidermidis</i> ATCC 12228 (WDCM 00036)	No or Atypical Growth	Pass
<i>Staphylococcus saprophyticus</i> ATCC 15305 (WDCM 00159)	No or Atypical Growth	Pass

This material complies with the 3M specifications for this product construction, and applicable criteria for routine quality control and microbiological performance of ISO 11133. 3M Wroclaw is certified to ISO 9001 through an independent agency.

3M Health Care
3M Center, Building 275-5W-05
St. Paul, MN 55144-1000 Phone: 1-800-328-1671

3M and Petrifilm are trademarks of 3M. Please recycle.
Printed in USA © 3M 2017. All rights reserved.

Version 11

STX
Staph Express Count Plate