

Clean-Trace® ATP Surface Swab Test


Item No. 700005215
Catalog No. UXC100

Kit Lot No: 3000039943
Date of Manufacture: 06 Dec 2025
Expiration Date: 06 Oct 2026

Test	Temp	Inst Serial Number	Specification	Test Result
Device Background	20°C	072	Mean Bk \leq 25 RLU	Pass
Device Activity	20°C	072	Activity Mean 6750 – 7570 RLU	Pass

Comments:

This is to certify that the product described above has been manufactured in accordance with Neogen specifications.



Barbara Lyczko
Date: 08 Dec 2025

Manufactured in the UK by: Neogen Food Safety UK Ltd
The Science Park
Bridgend CF31 3NA
Tel +44 (0) 1656 641400



620 Leshar Place • Lansing, MI 48912 • 800.234.5333 (U.S. and Canada) • 517.372.9200
The Dairy School • Auchincruive • Ayr, KA6 5HU • Scotland, UK • +44 (0) 1292 525 600