

## Clean-Trace® ATP Water Test


**Item No. 700005216**  
**Catalog No. H2O**

**Kit Lot No: 3000042172**  
**Date of Manufacture: 17 Nov 2025**  
**Expiration Date: 17 Nov 2026**

Test	Temp	Inst Serial Number	Specification	Test Result
Device Background	20°C	072	Mean Bk $\leq$ 20 RLU	Pass
Device Activity	20°C	072	Activity Mean 12150 – 16200 RLU	Pass

Comments:

This is to certify that the product described above has been manufactured in accordance with Neogen specifications.



Barbara Lyczko  
Date: 25 Nov 2025

Manufactured in the UK by: Neogen Food Safety UK Ltd  
The Science Park  
Bridgend CF31 3NA  
Tel +44 (0) 1656 641400



620 Leshar Place • Lansing, MI 48912 • 800.234.5333 (U.S. and Canada) • 517.372.9200  
The Dairy School • Auchincruive • Ayr, KA6 5HU • Scotland, UK • +44 (0) 1292 525 600