

# Certificate of Analysis



## DICHLORAN GLYCEROL (DG-18) AGAR BASE

**BULK Lot No: 30000446447**

**Exp. Date: JAN.31.2029**

**SKU: 700003197, 700003198, 700003199, 700003200, 700004469**

**No. NCM0081**

**This COA supersedes the one created 2/24/26**

### PHYSICAL CHARACTERISTICS

<u>SPECIFICATION</u>		<u>EXPECTED RESULTS</u>	<u>ACTUAL RESULTS</u>
<u>Dehydrated:</u>	Appearance	Homogenous, free-flowing powder	Meets Expected Result
	Color	Beige	Meets Expected Result
<u>Prepared:</u>	Clarity	Clear to Slightly Hazy	Meets Expected Result
	Color	Beige	Meets Expected Result
<u>pH:</u>		5.6 ± 0.2 @ 25°C	Meets Expected Result

### CULTURAL RESPONSE- tested in compliance to ISO 11133:2014/A2:2020

The medium was prepared according to label directions and inoculated with the organisms listed below. Cultures were incubated aerobically at 25 ± 1°C and examined for growth up to 5 days.

<u>MICROORGANISM</u>	<u>WDCM</u>	<u>APPROX. INOCULUM (CFU)</u>	<u>EXPECTED RESULTS</u>	<u>ACTUAL RESULTS</u>
<i>Saccharomyces cerevisiae</i>	00058	50-200	>50% Recovery	Meets Expected Result
<i>Walleimia sebi</i>	00182	50-200	>50% Recovery	Meets Expected Result
<i>Aspergillus restrictus</i>	00183	50-200	>50% Recovery	Meets Expected Result
<i>Eurotium rubrum</i>	00184	50-200	>50% Recovery	Meets Expected Result
<i>Escherichia coli</i>	00012	>10 <sup>4</sup>	No growth	Meets Expected Result
<i>Escherichia coli</i>	00013	>10 <sup>4</sup>	No growth	Meets Expected Result
<i>Bacillus subtilis</i>	00003	>10 <sup>4</sup>	Inhibited	Meets Expected Result

### ANIMAL DERIVED MATERIAL COUNTRY OF ORIGIN INFORMATION

<u>ANIMAL DERIVED MATERIAL</u>	<u>TISSUE SOURCE/ RISK CATEGORY*</u>	<u>ANIMAL SPECIES</u>	<u>COUNTRY OF ORIGIN</u>	<u>CERTIFICATE OF SUITABILITY</u>
Casein	Milk/B*	Bovine	<i>New Zealand</i>	<i>N/A</i>
Enzyme	Pancreas/ NA**	Porcine	<i>United States</i>	<i>N/A</i>

\* Risk Categories as defined by the European Pharmacopeia.

\*\* Not Applicable. "Non-TSE-relevant animal species."

Product was tested in compliance to the regulation noted above. Where a regulation is not referenced, Neogen Corporation standard test methods were observed. Variances in data may occur. External factors such as adverse storage conditions or contamination of product may cause loss in performance. Neogen Corporation certifies that this lot met all quality control specifications for this product.

  
Joe-Lyn McClain  
Quality Supervisor

Tuesday, March 10, 2026



620 Leshar Place • Lansing, MI 48912 • 800-234-5333 (USA/Canada) • 517-372-9200  
1 Quest Park • Heywood, BL9 7JJ, UK • +44 (0)161 820 3833  
foodsafety@neogen.com • foodsafety.neogen.com

CofA-400000792

Rev: 5