

CERTIFICATE OF ANALYSIS



**Petrifilm® Aerobic
Count Plates**
Item No. 700002116 or 700002120
Catalog No. 6400 or 6406

Lot No. 348673
Manufacture Date: 2026-02-03
Expiration Date: 2027-08-04

Organism Tested	Growth Specification	Result
<i>Lactococcus lactis</i> subsp. <i>cremoris</i> ATCC 19257	≥-3.0*	Pass
<i>Escherichia coli</i> ATCC 51813	≥-3.0*	Pass

*Expressed as the number of standard deviations away from the average count on standard agar medium.

ISO 11133:2014 Performance Testing

Organism Tested	Growth Specification	Result
<i>Escherichia coli</i> ATCC 25922 (WDCM 00013)	Productivity Ratio ≥ 0.7	Pass
<i>Bacillus subtilis</i> subsp. <i>spizizenii</i> ATCC 6633 (WDCM 00003)	Productivity Ratio ≥ 0.7	Pass
<i>Escherichia coli</i> ATCC 8739 (WDCM 00012)	Productivity Ratio ≥ 0.7	Pass
<i>Staphylococcus aureus</i> ATCC 25923 (WDCM 00034)	Productivity Ratio ≥ 0.7	Pass
<i>Staphylococcus aureus</i> ATCC 6538 (WDCM 00058)	Productivity Ratio ≥ 0.7	Pass

Chemical and Physical Testing

Parameter	Specification	Result	Pass
pH	7.0 ± 1.0	7.4	Yes

Contamination Check

Test Plan	Result
96 Randomized plates per batch tested per a statistical sampling plan	Pass

This material complies with the Neogen specifications for this product construction, and applicable criteria for routine quality control and microbiological performance. Neogen is certified to ISO 9001 through an independent agency.

Manufactured for Neogen Corporation by Solventum, 601 22nd Ave. South, Brookings, SD 57006

Gabrielle Jackson

Gabrielle Jackson
Date: 2026-03-02

CofA-700002116
or 700002120
FS-COA-0216
Rev: 8



620 Leshar Place • Lansing, MI 48912 • 800.234.5333 (U.S. and Canada) • 517.372.9200
The Dairy School • Auchincruive • Ayr, KA6 5HU • Scotland, UK • +44 (0) 1292 525 600