

CERTIFICATE OF ANALYSIS



Petrifilm® Yeast and Mold Count Plates

Item No. 700002121, 700002128 or
700006809

Catalog No. 6407, 6417 or 6445

Lot No. 348879

Manufacture Date: 2026-02-11

Expiration Date: 2027-08-12

Organism Tested	Minimum Growth	Result
<i>Aspergillus niger</i> , M6	Productivity Ratio \geq 0.7	Pass
<i>Paecilomyces</i> species, M10	Productivity Ratio \geq 0.7	Pass
<i>Botrytis species</i> , M97	Productivity Ratio \geq 0.7	Pass
<i>Hansenula anomala</i> , Y28	Productivity Ratio \geq 0.7	Pass
<i>Saccharomyces cerevisiae</i> , Y41	Productivity Ratio \geq 0.7	Pass
<i>Enterococcus faecalis</i>	No Growth	Pass
<i>Acinetobacter</i> species	No Growth	Pass

Chemical and Physical Testing

Parameter	Specification	Result	Pass
pH	7.0 \pm 1.0	7.06	Yes

Contamination Check

Test Plan	Result
96 Randomized plates per batch tested per a statistical sampling plan	Pass

This material complies with the Neogen specifications for this product construction, and applicable criteria for routine quality control and microbiological performance. Neogen is certified to ISO 9001 through an independent agency.

Manufactured for Neogen Corporation by Solventum, 601 22nd Ave. South, Brookings, SD 57006

Gabrielle Jackson

Gabrielle Jackson

Date: 2026-03-04