

CERTIFICATE OF ANALYSIS



**Petrifilm® *Bacillus cereus*
Count Plates**
Item No. 700006920 or 700006921
Catalog No. 6482 or 6483

Lot No. 418325157C

Manufacture Date: 2025-06-06

Expiration Date: 2026-12-05

ISO 11133:2014 Performance Testing

Organism Tested	Growth Specification	Result
<i>Bacillus cereus</i> ATCC 11778 (WDCM 00001)	Productivity Ratio ≥ 0.5	Pass
<i>Escherichia coli</i> ATCC 8739 (WDCM 00012)	No Growth	Pass
<i>Bacillus subtilis</i> subsp. <i>spizizenii</i> (WDCM 00003)	No or Atypical Growth	Pass

Chemical and Physical Testing

Parameter	Specification	Result	Pass
pH	7.0 \pm 1.0	7,66	Yes

Contamination Check

Test Plan	Result
96 Randomized plates per batch tested per a statistical sampling plan	Pass

This material complies with the Neogen specifications for this product construction, and applicable criteria for routine quality control and microbiological performance. Neogen is certified to ISO 9001 through an independent agency.

Manufactured for Neogen Corporation by 3M Poland, ul.Kowalska 143 | 51- 424 Wroclaw Poland

Agnieszka Grabowy

Date: Wednesday July 2, 2025

CofA-700006920 or
700006921
FS-COA- 1365
Rev: 2



620 Leshar Place • Lansing, MI 48912 • 800.234.5333 (U.S. and Canada) • 517.372.9200
The Dairy School • Auchincruive • Ayr, KA6 5HU • Scotland, UK • +44 (0) 1292 525 600