

CERTIFICATE OF ANALYSIS



Almond Protein Rapid Kit

Item No. 700002292

Catalog No. L25ALM

Kit Lot No. AZK2152B

Buffer Lot No. AZK2152B

Expiration Date: 2028-02-23

Performance Testing*

Negative Controls

| Sample | Specification | Result |
|----------------|---------------------|--------|
| Matrix Testing | Negative, ≥ 24 | Pass |

Positive Controls

| Sample | Specification | Result |
|----------------|---------------------|--------|
| Matrix Testing | Positive, ≥ 24 | Pass |

*A minimum of four matrices are used for performance testing. All results are interpreted by eye, holding the strip an arm's length away. If the control, hook and test lines are visible the test is considered positive. If only the control and hook lines are visible the test is considered negative. If either the control or hook lines aren't visible the test is considered invalid. Positive controls are spiked at 2ppm.

This letter certifies that the product described above has been tested and found to comply with Neogen's product performance specifications applicable to subject device at time of manufacture. Visit the Neogen Food Safety website, www.Neogen.com/foodsafety for additional information. Neogen is certified to ISO 9001 through an independent agency.

Manufactured for Neogen Corporation by VanPelt Biosciences: 18211-D Flower Hill Way, Gaithersburg, MD 20879

Christopher Norello

Date: 2026-02-23



CofA-700002292

FS-COA-0266

Rev: 3

620 Leshler Place • Lansing, MI 48912 • 800.234.5333 (U.S. and Canada) • 517.372.9200
The Dairy School • Auchincruive • Ayr, KA6 5HU • Scotland, UK • +44 (0) 1292 525 600