

CERTIFICATE OF ANALYSIS



Egg White Protein Rapid Kit
Item No. 700002290
Catalog No. L25EGG

Kit Lot No. EZM6214A
Buffer Lot No. EZM6214A
Expiration Date: 2028-01-22

Performance Testing*
Negative Controls

Sample	Specification	Result
Matrix Testing	Negative, ≥ 24	Pass

Positive Controls

Sample	Specification	Result
Matrix Testing	Positive, ≥ 24	Pass

*A minimum of four matrices are used for performance testing. All results are interpreted by eye, holding the strip an arm's length away. If the control, hook and test lines are visible the test is considered positive. If only the control and hook lines are visible the test is considered negative. If either the control or hook lines aren't visible the test is considered invalid. Positive controls are spiked at 2ppm.

This letter certifies that the product described above has been tested and found to comply with Neogen's product performance specifications applicable to subject device at time of manufacture. Visit the Neogen Food Safety website, www.Neogen.com/foodsafety for additional information. Neogen is certified to ISO 9001 through an independent agency.

Manufactured for Neogen Corporation by VanPelt Biosciences: 18211-D Flower Hill Way, Gaithersburg, MD 20879

Christopher Norello
Date: 2026-03-04

