

Clean-Trace® Water Plus - Total ATP Test


Item No. 700002100
Catalog No. AQT200

Kit Lot No: 3000045591
Date of Manufacture: 02 Feb 2026
Expiration Date: 02 Feb 2027

Test	Temperature	Inst Serial Number	Specification	Test Result
Device Background	20°C	072	Mean Bk \leq 20 RLU	Pass
Device Activity	20°C	072	Activity Mean 12150 – 16200 RLU	Pass

Comments:

This is to certify that the product described above has been manufactured in accordance with Neogen specifications.



Barbara Lyczko
Date: 11 Feb 2026

Manufactured in the UK by: Neogen Food Safety UK Ltd
The Science Park
Bridgend CF31 3NA
Tel +44 (0) 1656 641400



620 Leshar Place • Lansing, MI 48912 • 800.234.5333 (U.S. and Canada) • 517.372.9200
The Dairy School • Auchincruive • Ayr, KA6 5HU • Scotland, UK • +44 (0) 1292 525 600