

# CERTIFICATE OF ANALYSIS



## Soy Protein Rapid Kit

Item No. 700002296

Catalog No. L25SOY

Kit Lot No. SZM6934A

Buffer Lot No. SZM6934A

Expiration Date: 2028-02-23

### Performance Testing\*

#### Negative Controls

Sample	Specification	Result
Matrix Testing	Negative, $\geq 24$	Pass

#### Positive Controls

Sample	Specification	Result
Matrix Testing	Positive, $\geq 24$	Pass

\*A minimum of four matrices are used for performance testing. All results are interpreted by eye, holding the strip an arm's length away. If the control, hook and test lines are visible the test is considered positive. If only the control and hook lines are visible the test is considered negative. If either the control or hook lines aren't visible the test is considered invalid. Positive controls are spiked at 2ppm.

This letter certifies that the product described above has been tested and found to comply with Neogen's product performance specifications applicable to subject device at time of manufacture. Visit the Neogen Food Safety website, [www.Neogen.com/foodsafety](http://www.Neogen.com/foodsafety) for additional information. Neogen is certified to ISO 9001 through an independent agency.

Manufactured for Neogen Corporation by VanPelt Biosciences: 18211-D Flower Hill Way, Gaithersburg, MD 20879

Christopher Norello

Date: 2026-03-13



CofA-700002296

FS-COA-0277

Rev: 3

620 Leshler Place • Lansing, MI 48912 • 800.234.5333 (U.S. and Canada) • 517.372.9200

The Dairy School • Auchincruive • Ayr, KA6 5HU • Scotland, UK • +44 (0) 1292 525 600