

CERTIFICATE OF ANALYSIS



Petrifilm® Environmental Listeria Plates Item No. 700002233 or 700002232 Catalog No. 6447 or 6448

Lot No. 418325276A
Manufacture Date: 2025-10-03
Expiration Date: 2027-04-03

ISO 11133:2014 Performance Testing

Organism Tested	Growth Specification	Result
<i>Listeria monocytogenes</i> SLR-22	Typical Listeria Growth	Pass
<i>Listeria innocua</i> ATCC 33090	Typical Listeria Growth	Pass
<i>Enterococcus faecalis</i> ATCC 14506	No Typical Listeria Growth	Pass
<i>Bacillus atrophaeus (subtilis)</i> ATCC 9372	No Typical Listeria Growth	Pass

*Expressed as the number of standard deviations away from the average count on standard agar medium.

Chemical and Physical Testing

Parameter	Specification	Result	Pass
pH	7.0 ± 1.0	7,24	Yes

Contamination Check

Test Plan	Result
96 Randomized plates per batch tested per a statistical sampling plan	Pass

This material complies with the Neogen specifications for this product construction, and applicable criteria for routine quality control and microbiological performance. Neogen is certified to ISO 9001 through an independent agency.

Manufactured for Neogen Corporation by 3M Poland, ul.Kowalska 143 | 51- 424 Wroclaw Poland

Paulina Jabłońska

Date: Wednesday October 8, 2025

CofA-700002233

or 700002232

FS-COA-0298

Rev: 4



620 Leshler Place • Lansing, MI 48912 • 800.234.5333 (U.S. and Canada) • 517.372.9200
The Dairy School • Auchincruive • Ayr, KA6 5HU • Scotland, UK • +44 (0) 1292 525 600