CERTIFICATE OF ANALYSIS

Reveal Q+ MAX for Ochratoxin

SKU: 700002614
Product No.: 8688

Kit Lot No. 3000054839
Device Lot No. 127061
Exp. Date: 2027-04-30

PERFORMANCE

Incubation Time: **5 minutes**
Room Temperature: **~30°C**

Raptor

Raptor
Ochra Q+M
127061
Rev. 0

Raptor Solo

Solo
Ochra Q+M
127061
Rev. 0

READERS:

Raptor
Raptor Solo

TEST ITEM	SPECIFICATION	DECISION
Sensitivity:	Detects Ochratoxin levels at 5ppb - 25ppb. Detects Ochratoxin levels at 25ppb - 100ppb w/dilution.	Pass

Variances in data may occur. External factors such as stress from shipping, storage conditions, handling or contamination of components may cause loss of color activity. Neogen Corporation certifies that this test kit has met all internal quality control specifications for this product.

Quality Control Supervisor

Date: Tuesday, August 18, 2026

NEOGEN
NEOGEN.com
20 Lesher Place • Lansing, MI 48912 • 800.234.5333 (U.S. and Canada) • 517.372.9200
The Dairy School • Auchincruive • Ayr, KA6 5HU • Scotland, UK • +44 (0) 1292 525 600