# CERTIFICATE OF ANALYSIS

## *Salmonella* Enrichment Base, 2.5 kg

**Item No. 700002270**
**Catalog No. SEB025**

**Lot No.** 2605AR01
**Manufacture Date:** 2026-06-08
**Expiration Date:** 2029-06-07

**Chemical and Physical Testing**

| Parameter | Specification | Result | Pass |
|---|---|---|---|
| Appearance Color | Yellow to Amber | Yellow | Yes |
| Prepared Medium Color | Amber | Amber | Yes |
| Precipitate | None | None | Yes |
| pH | 7.0 ± 0.2 | 6.94 | Yes |

**Terminal Irradiation Processing Parameters**

This product has been subjected to Gamma irradiation as a terminal process. The minimum and maximum delivered met established specifications for this product.

This letter certifies that the product described above has been tested and found to comply with Neogen's product performance specifications applicable to subject device at time of manufacture. Visit the Neogen Food Safety website, www.Neogen.com/foodsafety for additional information. Neogen is certified to ISO 9001 through an independent agency.

Manufactured for Neogen Corporation by Alpha Biosciences, 3651 Clipper Mill Rd. Baltimore, MD 21211.

Gabrielle Jackson
Date: 2026-08-24



NEOGEN.com
620 Lesher Place • Lansing, MI 48912 • 800.234.5333 (U.S. and Canada) • 517.372.9200
The Dairy School • Auchincruive • Ayr, KA6 5HU • Scotland, UK • +44 (0) 1292 525 600