CERTIFICATE OF ANALYSIS

Clean-Trace® Surface ATP Swab Test

Item No. 700002207
Catalog No. UXL100

Kit Lot No: 3000057486
Date of Manufacture: 02 Oct 2026
Expiration Date: 02 Aug 2027

Test	Temp	Inst Serial Number	Specification	Test Result
Device Background	20°C	285	Mean Bk ≤ 25 RLU	Pass
Device Activity	20°C	285	Activity Mean 6750 – 7570 RLU	Pass

Comments:

This is to certify that the product described above has been manufactured in accordance with Neogen specifications.

Barbara Lyczko
Date: 05 Oct 2026

Manufactured in the UK by: Neogen Food Safety UK Ltd
The Science Park
Bridgend CF31 3NA
Tel +44 (0) 1656 641400