CERTIFICATE OF ANALYSIS

Petrifilm® Rapid Yeast and Mold Count Plates

Item No. 700002138 or 700002148
Catalog No. 6475 or 6477

Lot No. 30008888
Manufacture Date: 2026-07-22
Expiration Date: 2028-01-20

Organism Tested	Growth Specification	Result
Geotrichum candidum ATCC 34614	Productivity Ratio ≥ 0.5	Pass
Candida albicans ATCC 10231	Productivity Ratio ≥ 0.5	Pass
Enterococcus faecalis ATCC 14506	No Growth	Pass
Escherichia coli ATCC 25922	No Growth	Pass

Chemical and Physical Testing

Parameter	Specification	Result	Pass
pH	6.5 ± 1.0	6.10	Yes

Contamination Check

Test Plan	Result
96 Randomized plates per batch tested per a statistical sampling plan	Pass

This material complies with the Neogen specifications for this product construction, and applicable criteria for routine quality control and microbiological performance. Neogen is certified to ISO 9001 through an independent agency.

Manufactured for Neogen Corporation by Solventum, 601 22nd Ave. South, Brookings, SD 57006

Gabrielle Jackson

Gabrielle Jackson
Date: 2026-09-04

NEOGEN
NEOGEN.com
620 Lesher Place • Lansing, MI 48912 • 800.234.5333 (U.S. and Canada) • 517.372.9200
The Dairy School • Auchincruive • Ayr, KA6 5HU • Scotland, UK • +44 (0) 1292 525 600