CERTIFICATE OF ANALYSIS

Petrifilm® Yeast and Mold Count Plates

Item No. 700002121, 700002128 or 700006809
Catalog No. 6407, 6417 or 6445

Lot No. 30011687
Manufacture Date: 2026-07-25
Expiration Date: 2028-01-23

Organism Tested	Minimum Growth	Result
Aspergillus niger, M6	Productivity Ratio≥0.7	Pass
Paecilomyces species, M10	Productivity Ratio≥0.7	Pass
Botrytis species, M97	Productivity Ratio≥0.7	Pass
Hansenula anomala, Y28	Productivity Ratio≥0.7	Pass
Saccharomyces cerevisiae, Y41	Productivity Ratio≥0.7	Pass
Enterococcus faecalis	No Growth	Pass
Acinetobacter species	No Growth	Pass

Chemical and Physical Testing

Parameter	Specification	Result	Pass	Result
pH	7.0 ± 1.0	6.53	Yes	

Contamination Check

Test Plan	Result
96 Randomized plates per batch tested per a statistical sampling plan	Pass

This material complies with the Neogen specifications for this product construction, and applicable criteria for routine quality control and microbiological performance. Neogen is certified to ISO 9001 through an independent agency.

Manufactured for Neogen Corporation by Solventum, 601 22nd Ave. South, Brookings, SD 57006

Gabrielle Jackson

Gabrielle Jackson
Date: 2026-09-10

NEOGEN
NEOGEN.com
620 Lesher Place • Lansing, MI 48912 • 800.234.5333 (U.S. and Canada) • 517.372.9200
The Dairy School • Auchincruive • Ayr, KA6 5HU • Scotland, UK • +44 (0) 1292 525 600