CERTIFICATE OF ANALYSIS

Petrifilm® Environmental *Listeria* Plates

Item No. 700002233 or 700002232
Catalog No. 6447 or 6448

Lot No. 418326193A
Manufacture Date: 2026-07-12
Expiration Date: 2028-01-10

ISO 11133:2014 Performance Testing

Organism Tested	Growth Specification	Result
Listeria monocytogenes SLR-22	Typical Listeria Growth	Pass
Listeria innocua ATCC 33090	Typical Listeria Growth	Pass
Enterococcus faecalis ATCC 14506	No Typical Listeria Growth	Pass
Bacillus atrophaeus (subtilis) ATCC 9372	No Typical Listeria Growth	Pass

*Expressed as the number of standard deviations away from the average count on standard agar medium.

Chemical and Physical Testing

Parameter	Specification	Result	Pass
pH	7.0 ± 1.0	7.19	Yes

Contamination Check

Test Plan	Result
96 Randomized plates per batch tested per a statistical sampling plan	Pass

This material complies with the Neogen specifications for this product construction, and applicable criteria for routine quality control and microbiological performance. Neogen is certified to ISO 9001 through an independent agency.

Manufactured for Neogen Corporation by 3M Poland, ul.Kowalska 143 | 51- 424 Wroclaw Poland

Jarosław Matczak

Date: Thursday July 16, 2026

NEOGEN.com
620 Lesher Place • Lansing, MI 48912 • 800.234.5333 (U.S. and Canada) • 517.372.9200
The Dairy School • Auchincruive • Ayr, KA6 5HU • Scotland, UK • +44 (0) 1292 525 600