CERTIFICATE OF ANALYSIS

Salmonella Enrichment Supplement, 1 gram

Item No. 700002097
Catalog No. SESUP001

Lot No. R012-144
Expiration Date: 2028-07-29

Chemical and Physical Testing

Parameter	Specification	Result	Pass
Powdered Medium Color	Light Green	Light Green	Yes
Prepared Medium Color	Blue to Green	Blue to Green	Yes
Precipitate	None to Trace	None to Trace	Yes
pH	7.0 ± 0.2	7.07	Yes

Microorganism Testing

Organism Tested	Specification	Result	Pass
Salmonella enterica Typhimurium ATCC 14028	Visible Turbidity	Visible Turbidity	Yes
Salmonella enteritidis ATCC 49223	Visible Turbidity	Visible Turbidity	Yes
Citrobacter freundii 548	No Visible Turbidity	No Visible Turbidity	Yes
Escherichia coli 723	No Visible Turbidity	No Visible Turbidity	Yes

Terminal Irradiation Processing Parameters

This product has been subjected to Gamma irradiation as a terminal process. The minimum and maximum delivered met established specifications for this product.

This letter certifies that the product described above has been tested and found to comply with Neogen's product performance specifications applicable to subject device at time of manufacture. Visit the Neogen Food Safety website, www.Neogen.com/foodsafety for additional information. Neogen is certified to ISO 9001 through an independent agency.

Manufactured for Neogen Corporation by TOKU-E: 4112 Bakerview Spur, Bellingham, WA 98226

Kevin Bender
Date: 2026-08-17

NEOGEN.com
620 Lesher Place • Lansing, MI 48912 • 800.234.5333 (U.S. and Canada) • 517.372.9200
The Dairy School • Auchincruive • Ayr, KA6 5HU • Scotland, UK • +44 (0) 1292 525 600